CERTIFICATE OF QUALITY

Filtration Kit for Growth Cassette

Part No: B150-048 Lot No: 105036 Expiry: 31JAN22

Storage: Room temperature

PRODUCT DESCRIPTION:

The Filtration Kit consists of a black 0.45 micron pore size cellulose ester membrane held in a plastic base below a removable filtration funnel.

MANUFACTURE:

The product was manufactured in an ISO qualified facility. The Gamma treatment was performed in an ISO qualified facility.

QUALITY RELEASE CRITERIA:

The quality control data given in this document represent the quality of the released lot as tested.

PHYSICAL TESTS:

1. Flow rate	Pass
2. Baseline signal	Pass

BIOLOGICAL TESTS:

Growth Promotion:

Organisms from the following panel were shown to be retained by the membrane and give good growth when placed on qualified media.

Bacillus subtilis	[1]ATCC® 6633	Pass
Psuedomonas aeruginosa	[1]ATCC® 9027	Pass
Staphylococcus aureus	[1]ATCC® 6538	Pass

[1]Or equivalent

This is Gamma Radiation treated product, packaged as 8 sets of 6 units in tyvek/poly sleeves stored in a cardboard box.

Released by:
Quality Assurance

Date: 02 Sep 2020

Rapidmicro biosystems

1001 Pawtucket Blvd.
Lowell, MA 01854
978-349-3200

This product is not for In Vitro diagnostic use.
Country of origin- USA