Rapid Micro Biosystems

Certificate of Analysis

Product Name:	Growth Direct Vision Lids		
Catalogue numbers:	☒ EMVL-100	**Manufacture Date:**	16Aug2022
	☐ BBVL-048		
Lot Number:	133022	**Expiration Date:**	31Aug2024

The above lot has been produced in an ISO9001 certified facility and treated with ethylene oxide in accordance with Rapid Micro Biosystems Quality System. Representative samples were tested per Rapid Micro Biosystems protocols and were found to meet the performance and physical characteristics release criteria for this product.

Storage: Room Temperature

Product Description:
The Growth Direct Vision Lid is a transparent non-fluorescing plastic lid that allows images to be taken by the Growth Direct of the membrane on the enclosed media surface.

Quality Release Criteria:
The quality control data given in this document represent the quality of the released lot as tested.

Test Performed	Test Result
Optical Background	Pass
No fluorescent debris	Pass

Ethylene Oxide Treatment:

Parameter	Specification	Test Result
EO Inject Pressure	8.4 – 8.8 PSIA	Pass
EO Inject Time	5 – 45 min	Pass
EO Inject Temperature	0 – 135 °F	Pass
EO Dwell Pressure	14.0 – 14.5 PSIA	Pass
EO Dwell Time	240 – 250 min	Pass
EO Dwell Temperature	125 – 135 °F	Pass

Certificate results reported here were obtained at time of release. Product is for exclusive use with Rapid Micro Biosystems Growth Direct system. Not for In Vitro Diagnostic use.

Released by: _____ Release Date 22Sep2022
Quality Representative

Rapid Micro Biosystems, Inc.
1001 Pawtucket Blvd West, Suite 280, Lowell, MA 01854
Phone: +1 978-349-3200 Fax: +1 978-349-2065
www.rapidmicrobio.com
Email:support@rapidmicrobio.com