Rapid Micro Biosystems

Certificate of Analysis

Product Name:	**Bioburden Filtration Funnel**		
Catalogue numbers:	☐ B150-048 ☒ B250-048	**Manufacture Date:**	**24Apr2023**
Lot Number:	**146086**	**Expiration Date:**	**31Oct2024**

The above lot has been produced in an ISO9001-certified facility and sterilized in accordance with Rapid Micro Biosystems Quality Systems. Representative samples were tested per Rapid Micro Biosystems protocols and were found to meet the performance and physical characteristics release criteria for this product.

Storage: Room temperature

Product Description:

The Filtration Kit consists of a black 0.45 micron pore size cellulose ester membrane held in a plastic base below a removable filtration funnel.

Quality Release Criteria:

The quality control data given in this document represent the quality of the released lot as tested.

Physical Test Performed	Test Result
Flow Rate	Pass
Baseline Signal	Pass
Growth Promotion	**Test Result**
Bacillus subtilis - ATCC® 6633[1]	Pass
Pseudomonas aeruginosa - ATCC® 9027[1]	Pass
Staphylococcus aureus - ATCC® 6538[1]	Pass

[1]Or equivalent

Irradiation:

This is Gamma Radiation treated product at 15-40kGy, packaged as 8 sets of 6 units in poly sleeves stored in a cardboard box.

Certificate results reported here were obtained at time of release. Product is for exclusive use with Rapid Micro Biosystems Growth Direct system. Not for In Vitro Diagnostic use.

Released by: _____ Release Date 06JUN2023

Quality Representative

Rapid Micro Biosystems, Inc.
1001 Pawtucket Blvd West, Suite 280, Lowell, MA 01854
Phone: +1 978-349-3200 Fax: +1 978-349-2065
www.rapidmicrobio.com
Email:support@rapidmicrobio.com